

## Pharmacy\_Subscriberlist@DCA

---

**From:** General Board of Pharmacy Subscriber List <PHARM-GENERAL@DCALISTS.CA.GOV>  
on behalf of Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>  
**Sent:** Monday, February 08, 2016 1:54 PM  
**To:** PHARM-GENERAL@DCALISTS.CA.GOV  
**Subject:** Product Recall - AstraZeneca

AstraZeneca is recalling the item lots below due to the counter ring may not be initially set to "60" in the viewing window. The device may display a blank window or a number below 60. If a blank window is present upon opening, there is a potential for the device to become empty before the counter gets to 0 and locks out, resulting in no dose being delivered.

TUDORZA PRESSAIR Metered Dose Inhaler (aclidinium bromide inhalation powder) 400 mcg/actuation. 60 per carton

NSC: 0310-0800-60

Lot Number	Expiration	Ship Dates
1144394	4/2018	7/1/2015 to 9/15/2015
1145539	4/2018	7/23/2015 to 9/1/2015
1145868	5/2018	8/11/2015 to 9/24/2015

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

----- To unsubscribe from this email list please click on the link below and follow the instructions on the web page.

<https://www.dca.ca.gov/webapps/pharmacy/subscribe.php>