
From: Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>
Sent: Thursday, April 27, 2017 2:39 PM
To: PHARM-GENERAL@DCALISTS.CA.GOV
Cc: Pharmacy_Subscriberlist@DCA
Subject: Recall notice - B Braun

B Braun is recalling the item/lots below due to the potential for some blisters to be punctured resulting in a compromised sterility barrier. This recall is to the pharmacy level. Affected product started shipping November 7, 2012.

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

Description	Lot #	Exp Date	NDC	UPC
CYTOGUARD	1570315301	01/04/20	08021456082	12045608222
CLOSE MALE	1470311101	12/02/19		
CONCT CAP	1370273801	09/01/18		
	1370270401	08/05/18		
	1370268401	07/14/18		
	1370259201	05/09/18		
	1370257101	04/08/18		
	1370251101	03/06/18		
	1370249001	01/13/18		
	1240156004	09/19/17		
	1240156003	09/19/17		
	1240156002	09/19/17		
	1240156001	09/10/17		