

## Pharmacy\_Subscriberlist@DCA

---

**From:** Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>  
**Sent:** Wednesday, December 23, 2015 8:58 AM  
**To:** PHARM-GENERAL@DCALISTS.CA.GOV  
**Cc:** Pharmacy\_Subscriberlist@DCA  
**Subject:** Recall notice - Baxter Healthcare Corporation

Baxter Healthcare Corporation is recalling the items below for lack of evidence supporting the packaging claim for 12-month shelf life on released product. This recall is to the Healthcare Professional, Consumer level. Affected product started shipping November 20, 2014.

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

Description	Lot #	Exp Date	NDC	UPC
RAPID-FILL SYR STRIP WHT CT200	ALL	ALL	17191090200	31719190200
RAPID-FILL SYR STRIP YEL CT200	ALL	ALL	17191090221	31719190221
RAPID-FILL SYR STRIP ORG CT200	ALL	ALL	17191090222	31719190222
RAPID-FILL SYR STRIP RED CT200	ALL	ALL	17191090223	31719190223
RAPID-FILL SYR STRIP BLU CT200	ALL	ALL	17191090225	31719190225
RAPID-FILL SYR STRIP SAL CT200	ALL	ALL	17191090227	31719190227
RAPID-FILL SYR STRIP VIO CT200	ALL	ALL	17191090229	31719190229
RAPID-FILL SYR STRIP GRY CT200	ALL	ALL	17191090231	31719190231
RAPID-FILL SYR STRIP GRN CT200	ALL	ALL	17191090232	31719190232