From: General Board of Pharmacy Subscriber List on behalf of Board of Pharmacy
To: PHARM-GENERAL@DCALISTS.CA.GOV
Subject: Product Recall - Baxter Biosurgery
Date: Monday, May 23, 2016 2:15:00 PM

Baxter Biosurgery is recalling the item(s)/lot(s) below due to the potential of incomplete dissolution of polyethylene glycol (PEG) component during reconstitution of the product which may affect the consistency of the hydrogel formation during use. This recall is to the Consumer level. Affected product started shipping November 10, 2015.

Description	Lot #	Exp Date	UPC
MPB COSEAL 934070 2ML BAX	HA160136	01/31/17	99999934070
MPB COSEAL 934071 4ML BAX	HA151026	09/30/16	99999934071
	HA151035	10/31/16	
	HA151036	10/31/16	
	HA151205	11/30/16	
	HA160151	01/31/17	
	HA160229	02/28/17	
MPB COSEAL 934072 8ML BAX	HA151028	09/30/16	99999934072
	HA151037	10/31/16	
	HA151038	10/31/16	
	HA151220	12/31/16	
	HA160114	12/31/16	
	HA160115	01/31/17	
	HA160222	02/28/17	

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

---------- To unsubscribe from this email list please click on the link below and follow the instructions on the web page.

https://www.dca.ca.gov/webapps/pharmacy/subscribe.php