

---

**From:** Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>  
**Sent:** Thursday, June 01, 2017 8:57 AM  
**To:** PHARM-GENERAL@DCALISTS.CA.GOV  
**Cc:** Pharmacy\_Subscriberlist@DCA  
**Subject:** Recall notice - BD

BD is conducting a product removal recall of lot 6152995 of the BD PrecisionGlide™ Needle 18G x 1" RB , Cat (Ref) 305195, due to hub damage resulting in breakage and/or leakage during use. BD distributed the affected recalled lot from September 9, 2016 to October 31, 2016.

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

Product Name	Catalog (Ref) No.	Lot No.
BD PrecisionGlide™ Needle 18G x 1" RB	305195	6152995