
From: Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>
Sent: Thursday, September 07, 2017 7:58 AM
To: PHARM-GENERAL@DCALISTS.CA.GOV
Cc: Pharmacy_Subscriberlist@DCA
Subject: Recall notice - BD

In May 25, 2017, BD initiated a product recall for Catalog (Ref) 328466, Lot 6291768, since some polybags in the lot were incorrectly labeled as BD Ultra-Fine needle ½mL 8mm 31G, Cat (Ref) 328468. After further investigation, BD identified that two additional lots were also mislabeled with the same condition. As a result, BD is expanding the product removal recall of the BD Insulin Syringes with the BD Ultra-Fine needle ½mL 12.7mm 30G, Cat (Ref) 328466, to include all lots listed in the table below. Using a 12.7mm needle for injection when an 8mm was intended, could result in an increased risk of an inadvertent intramuscular injection, which may lead to unanticipated hypoglycemia.

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

Product Name	Catalog (Ref) No.	NDC/HRI # in Shelf Box	NDC/HRI # in Polybag	Lot No.
BD Insulin Syringes with the BD Ultra-Fine needle ½ mL 12.7mm 30G	328466	08290-3284-66	08290-8466-01	6291768
				6312558
				6340590