

---

**From:** Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>  
**Sent:** Monday, June 12, 2017 1:24 PM  
**To:** PHARM-GENERAL@DCALISTS.CA.GOV  
**Cc:** Pharmacy\_Subscriberlist@DCA  
**Subject:** Product Recall - Becton Dickinson

Becton Dickinson is recalling the items/lots below due to the product being potentially mislabeled. This recall is to the Consumer level. Affected product started shipping March 3, 2017.

Description	Lot #	Exp Date	NDC	UPC
SYR INS BD 0.5ML 30GX1/2 100	6291768	n/a	08290328466	3829
.5 mL 12.7 mm 30G BD Insulin Syringe	6291768	n/a	08290846601	&nb
.5 mL 8 mm 31G BD Insulin Syringe	6291768	n/a	08290846801	&nb

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.