

From: Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>
Sent: Thursday, March 16, 2017 8:30 AM
To: PHARM-GENERAL@DCALISTS.CA.GOV
Cc: Pharmacy_Subscriberlist@DCA
Subject: Recall notice - Medtronic

Medtronic is recalling specific item codes and production lots of Covidien Curity™ eye pad, Curity™ eye pad oval, Kerlix™ super sponge saline dressing, Curity™ wet dressing, Curity™ sodium chloride dressing and Curity™ saline dressing. This recall is being conducted due to the potential for the sterile packaging to be compromised. The use of products with this condition may result in a potentially increased risk for infection.

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

Item Code	Item Description	Lot Number beginning with	Expiration Date
03201	Covidien Curity™ Eye Pad	12, 13, 14, 15, 16	From 2017-02 through 2021-11
2841	Covidien Curity™ Eye Pad Oval		
91650	Covidien Curity™ Eye Pad		
3337	Covidien Curity™ Wet Dressing	14, 15, 16 *Exclude lots 16J098062 16J098162 16J098262	From 2017-02 through 2019-11
3338	Covidien Kerlix™ Super Sponge Saline Dressing		
3339	Covidien Curity™ Sodium Chloride Dressing		
3606*	Covidien Curity™ Saline Dressing		