

---

**From:** Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>  
**Sent:** Monday, April 03, 2017 4:43 PM  
**To:** PHARM-GENERAL@DCALISTS.CA.GOV  
**Cc:** Pharmacy\_Subscriberlist@DCA  
**Subject:** Recall Notice - Mylan

Mylan is recalling the items/lots below due to the result of two previously disclosed reports outside of the U.S. of failure to activate the device due to a potential defect in a supplier component. This recall is to the Consumer level. Affected product started shipping December 2015.

Description	Lot #	Exp Date	NDC	UPC
EPIPEN 0.3MG 0.3ML 1:1M 2PK	6GM199 10/31/17; 6GM198 10/31/17;			
	6GM091 10/31/17; 6GM088 10/31/17;			
	6GM087 10/31/17; 6GM082 09/30/17;			
	6GM081 09/30/17; 6GM072 09/30/17;			
EPIPEN JR .15MG .3ML 1:2M 2PK	5GM640 05/31/17; 5GM631 04/30/17;		49502050002	34950250002
	6GN215	09/30/17		
	5GN773	04/30/17		
	5GN767	04/30/17	49502050102	34950250102

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.