

From: [General Board of Pharmacy Subscriber List](#) on behalf of [Board of Pharmacy](#)
To: PHARM-GENERAL@DCALISTS.CA.GOV
Subject: Product Recall - Teleflex Medical
Date: Friday, November 04, 2016 1:26:36 PM

Teleflex Medical is recalling the item(s)/lot(s) below due to the potential that they may not deliver a fully atomized plume of medication. This recall is to the pharmacy level. Affected product started shipping January 2016.

Description	Lot #	NDC	UPC
LMA MAD 100 NASAL	160523; 160813; 160802; 160707; 160105; 160609; 160506; 160435; 160402; 160321; 160302; 160137; 160620;		
LMA MAD 100OS NASAL	160630; 160322; 160524;		
LMA MAD 110 NASAL	160507; 160217;		
LMA MAD 110OS NASAL	160240; 160312;		
LMA MAD 130 NASAL	160107; 160138; 160517;		
LMA MAD 130OS NASAL	160436; 160803;		
LMA MAD 140 NASAL	160218; 160437; 160610; 160801; 160125;		
LMA MAD 140OS NASAL	160438; 160226; 160727;		
LMA MAD 300 NASAL 25	160718; 160518; 160602; 160611; 160621; 160631; 160108; 160708; 160432; 160728; 160800; 160804; 160814; 160816; 160823; 160701; 160300; 160117; 160126; 160145; 160146; 160200; 160219; 160500; 160231; 160440; 160313; 160327; 160400; 160409; 160422; 160225;	60112031306	6011231306
LMA MAD 300B NASAL	160410		

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

----- To
 unsubscribe from this email list please click on the link below and follow the
 instructions on the web page.

<https://www.dca.ca.gov/webapps/pharmacy/subscribe.php>