

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

**AMENDED TO INCLUDE ADDITIONAL MATERIAL NUMBER (IN BOLD)**

Description	Lot #	Exp Date	NDC	UPC
SOD CHL 0.9% BG PVCFR 250ML24	J1H138	06/30/2023	00264780020	30264780020
	J1H137	06/30/2023		
	J1E213	05/31/2023		
	J1E204	05/31/2023		
	J1E086	05/31/2023		
<b>SOD CHL 0.9% BG 250ML BRAU 24</b>	J1H138	06/30/2023	<b>264780020</b>	<b>30264780020</b>
	J1H137	06/30/2023		
	J1E213	05/31/2023		
	J1E204	05/31/2023		
	J1E086	05/31/2023		
VA SOD CHL 0.9% BG250ML TRIL24	J1H138	06/30/2023	00264780020	30264780020
	J1H137	06/30/2023		
	J1E213	05/31/2023		
	J1E204	05/31/2023		
	J1E086	05/31/2023		

B. Braun is recalling the above item/lot(s) due to the potential for fluid leakage or low fill volume of the respective containers. There is also a potential for smaller micro-leaks. This recall is to the consumer level. Affected product started shipping June 2, 2021.