

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

| Description | Lot # | NDC | UPC |
|--------------------------------|--------|-------------|------------|
| VA TRUE METRIX SLF MONITOR MTR | KW0135 | 56151147401 | 2129200606 |

Trividia is voluntarily recalling the above item/lot due to incorrect factory-set unit of measure. This recall is to the consumer level. Affected product started shipping February 2019. **NOTE: This recall only affects the noted lot for one specific serial number: TA1548753. The serial number will be printed on the outer case/shipper label, or on the side of the individual meter box.**