

*The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.*

Medline Industries, LP is issuing a recall for specific item(s) and lot(s) of SwabFlush Prefilled Syringes with ICU Medical SwabCaps. ICU Medical has issued a recall of their SwabCap™ Disinfecting Cap for Needle-free Connectors, an integrated component of the SwabFlush syringe, after identifying a lot-specific issue where the seal between the foil lid and the plastic container of the SwabCap may be incomplete due to a manufacturing issue. This may result in evaporation of the isopropyl alcohol on the sponge. If the defect is not identified during setup and a defective SwabCap is used, this may result in a reduced level of disinfection of the luer access valve, potentially resulting in patient infection/sepsis. To date, ICU Medical has received zero (0) reports of serious injury or deaths associated with this issue. Affected SwabFlush Prefilled Syringes with ICU Medical SwabCaps are to be destroyed. Please additionally review your inventory for impacted SwabCaps (lots 13887153, 13914602, 13914610, and 13937147) that may have been removed from the SwabFlush syringe and reserved for later use.

NDC	Medline Item	Lot Numbers
		3144609
		3144857
63807010301	EMZE010301	3144886
		3144967
		3145090
		3145186