

*The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.*

Glenmark Specialty SA, Switzerland, Glenmark Pharmaceuticals Inc., USA is issuing a recall at Retail level for the following Ryaltris Nasal Spray batches as distributed by Hikma Specialty USA, Inc., Columbus, OH 43228. This recall is being conducted with the knowledge of the FDA.

The Recall of these Ryaltris Nasal Spray batches has been initiated due to a market complaint for “Spray not working” as applicable to Batch# 14230425, Expiry:11/2025.

The cause of this complaint as confirmed by the pump supplier, Aptar Pharma, is blockage of the dip tube caused by small particles of raw material that were created during the manufacturing process of the dip tube. The vendor lot number, 0038275600, was used in the production of the seven (7) batches of Ryaltris Nasal Spray supplied to U.S. market as identified in the table below.

<b>Sr. No.</b>	<b>NDC</b>	<b>Pack Size</b>	<b>Batch Number</b>	<b>Expiry Date</b>
1	59467-700-27	240 MD	14230425	Nov-25
2	59467-700-27	240 MD	14240024	Dec-25
3	59467-700-27	240 MD	14240029	Dec-25
4	59467-700-27	240 MD	14240076	Jan-26
5	59467-700-27	240 MD	14240082	Jan-26
6	59467-700-27	240 MD	14240090	Jan-26
7	59467-700-27	240 MD	14240100	Jan-26