

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

BD Medication Delivery Solutions is issuing a recall for BD Luer Tip Caps. BD has confirmed through internal testing that certain BD Luer Tip Cap trays failed routine biocompatibility testing.

Although the tip cap covers a small surface area of the syringe and does not come in direct contact with patients there is the potential that the failure of a biocompatibility test may have varying consequences ranging from no harm or clinical impact to potential for sensitization, irritation, pyrogenicity, acute systemic toxicity or hemolysis depending on the nature of and duration of contact with the syringe.

For clinicians who may have already used the affected syringe tip cap trays, use discretion in managing patients who may have been impacted.

To date, there have been no reported complaints or adverse events related to this issue.

Product Name	Catalog No.	Lot No.	Expiration Date	UDI-DI	Product Package Size
BD Luer Tip Caps	308341	5029670	31Dec2029	(01)50382903083416	10 caps per tray/ 200 trays per case
		5034211	31Jan2030		
		5041975	31Jan2030		