

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

Convatec is recalling a batch of AQUACEL® Foam Ag Adhesive 10cm x 10cm (x 10 pack) due to the identification of visual contamination in the form of brown spots present on several dressings.

REF Number	Batch	Description
420681	IC03858	AQUACEL® Foam Ag Adhesive 10cm x 10cm (x 10 pack)