

The Board of Pharmacy has received notice of the following product safety alert:

Description	Lot #	Exp Date	NDC	UPC
EXACTAMED O/DSP 5ML AMB8105 1C	B304N045P	02/18/2029	17191008105	31719108105
	B302M338P	12/09/2028		
	B301N070P	03/17/2029		
	B301N017P	01/22/2029		
EXACTAMED O/DSP 5ML AMB8505 5C	B302N077P	03/19/2029	17191008505	31719108505
	B301N049P	02/22/2029		
	B301N023P	01/28/2029		
EXACTAMED O/DSP 5ML Baxter Code: H93884205	B307N015P			
	B303N042P			
	B302M333P			
	B301N056P			

Baxter is issuing a safety alert for the above items/lots due to the potential for volume of a 10mL dispenser to be misinterpreted or not recognized based on the package labeling. This could result in medication errors. This safety alert is to the retail level. Affected product started shipping January 29, 2019.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.