

The Board of Pharmacy has received notice of the following product recall:

LEVEL OF NOTIFICATION: Consumer

SUPPLIER: KT Health

Description	Lot #	NDC	UPC
KT ICE/HEAT MASSG BALL S/BLKDS	All Lots	N/A	81417902222
KT RECOVERY+ ICE/HEAT MASSAGE BALL, 100000693	All Lots	N/A	81417902272
KT RECOVERY+ ICE/HEAT MASSAGE BALL, 10001606	All Lots	N/A	81417902256
KT RECOVERY+ ICE/HEAT MASSAGE BALL, 10001677	All Lots	N/A	81417902256

KT Health is recalling the above items due to reports of product failure resulting in the release of hot liquid from the device. This recall is to the consumer level. Affected product started shipping September 2017.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.