

---

**From:** Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>  
**Sent:** Wednesday, December 06, 2017 1:56 PM  
**To:** PHARM-GENERAL@DCALISTS.CA.GOV  
**Cc:** Pharmacy\_Subscriberlist@DCA  
**Subject:** Recall notice - Mallinckrodt

Mallinckrodt is recalling the item/lots below due to the filter in the insufflation filter assembly not meeting claims made in the device's 510(k) premarket notification. This recall is to the pharmacy level.

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

| Description          | Lot # | NDC         | UPC         |
|----------------------|-------|-------------|-------------|
| RAPLIXA DELIVERY KIT | ALL   |             |             |
| 1                    | LOTS  | 18868848161 | 31886884816 |