

The Board of Pharmacy has received notice of the following product recall:

PTS Diagnostics is voluntarily recalling lot number P901 of the PTS Panels® Lipid Panel test strips due to the potential for mislabeling of the MEMo Chip® included with the product.

Product Name: Lipid Panel Test Strips

REF Number: 1710

Lot Number P901

Expiration Date (month/date/year): 08/17/2020

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.