

The Board of Pharmacy has received notice of the following product correction:

**LEVEL OF NOTIFICATION:** Retail

**SUPPLIER:** Roche Diabetes

<b>Description</b>	<b>Lot #</b>	<b>NDC</b>	<b>UPC</b>
ACCU-CHEK GUIDE METER	92311000001	65702072910	36570272910

Roche Diabetes is issuing an Urgent Medical Device Correction for the above item due to battery life, battery life notification defects, and device power-on issues. This Urgent Medical Device Correction is to the retail level.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.