
From: Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>
Sent: Thursday, December 14, 2017 3:22 PM
To: PHARM-GENERAL@DCALISTS.CA.GOV
Cc: Pharmacy_Subscriberlist@DCA
Subject: Recall notice - Simple Diagnostics

Simple Diagnostics, is recalling three lots of Pharmacist Choice alcohol prep pads, which were manufactured by Foshan Flying Medical Products Co. Ltd., located in China, due to the lack of sterility assurance and other quality issues. The use of impacted alcohol prep pads could result in adverse events such as infections.

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

Product Description	NDC	Lot #	Distribution Dates
UPC		Expiration Date	
Pharmacist Choice Alcohol Prep Pads	98302-0001-05	SD2070421201	10/18/2016 – 07/19/2017
		12/2019	
UPC 898302001050		SD2070420925	
		09/2019	
		SD2070420601	
		12/2019	